

Quality Control for the BIOFIRE® BCID2 Panel¹

350765-1
2022-02

Rx Only

2022-11-16

Once opened, the product may be used until expiration date².

BIOFIRE FilmArray® Torch or 2.0 System

Antimicrobial Resistance Genes ³		
Gene ⁴	LOT 21711150 Control ⁵ 1-GN	LOT 21711151 Control ⁵ 2-GPY
CTX-M	Detected ⁹	Not Detected ¹⁰
IMP	Detected ⁹	Not Detected ¹⁰
KPC	Detected ⁹	Not Detected ¹⁰
<i>mcr-1</i>	Detected ⁹	Not Detected ¹⁰
<i>mecA/C</i>	Not Detected ¹⁰	Detected ⁹
<i>mecA/C</i> and MREJ (MRSA)	Not Detected ¹⁰	Detected ⁹
NDM	Detected ⁹	Not Detected ¹⁰
OXA-48-like	Detected ⁹	Not Detected ¹⁰
<i>vanA/B</i>	Not Detected ¹⁰	Detected ⁹
VIM	Detected ⁹	Not Detected ¹⁰

Gram-Positive Bacteria ⁴		
Pathogen ⁷	LOT 21711150 Control ⁵ 1-GN	LOT 21711151 Control ⁵ 2-GPY
<i>Enterococcus faecalis</i>	Not Detected ¹⁰	Detected ⁹
<i>Enterococcus faecium</i>	Not Detected ¹⁰	Detected ⁹
<i>Listeria monocytogenes</i>	Not Detected ¹⁰	Detected ⁹
<i>Staphylococcus</i> spp.	Not Detected ¹⁰	Detected ⁹
<i>Staphylococcus aureus</i>	Not Detected ¹⁰	Detected ⁹
<i>Staphylococcus epidermidis</i>	Not Detected ¹⁰	Detected ⁹
<i>Staphylococcus lugdunensis</i>	Not Detected ¹⁰	Detected ⁹
<i>Streptococcus</i> spp.	Not Detected ¹⁰	Detected ⁹
<i>Streptococcus agalactiae</i> (Group B)	Not Detected ¹⁰	Detected ⁹
<i>Streptococcus pneumoniae</i>	Not Detected ¹⁰	Detected ⁹
<i>Streptococcus pyogenes</i> (Group A)	Not Detected ¹⁰	Detected ⁹

Gram-Negative Bacteria ⁴		
Pathogen ⁷	LOT 21711150 Control ⁵ 1-GN	LOT 21711151 Control ⁵ 2-GPY
<i>Acinetobacter calcoaceticus-baumannii</i> complex	Detected ⁹	Not Detected ¹⁰
<i>Bacteroides fragilis</i>	Detected ⁹	Not Detected ¹⁰
<i>Enterobacterales</i>	Detected ⁹	Not Detected ¹⁰
<i>Enterobacter cloacae</i> complex	Detected ⁹	Not Detected ¹⁰
<i>Escherichia coli</i>	Detected ⁹	Not Detected ¹⁰
<i>Klebsiella aerogenes</i>	Detected ⁹	Not Detected ¹⁰
<i>Klebsiella oxytoca</i>	Detected ⁹	Not Detected ¹⁰
<i>Klebsiella pneumoniae</i> group	Detected ⁹	Not Detected ¹⁰
<i>Proteus</i> spp.	Detected ⁹	Not Detected ¹⁰
<i>Salmonella</i> spp.	Detected ⁹	Not Detected ¹⁰
<i>Serratia marcescens</i>	Detected ⁹	Not Detected ¹⁰
<i>Haemophilus influenzae</i>	Detected ⁹	Not Detected ¹⁰
<i>Neisseria meningitidis</i>	Detected ⁹	Not Detected ¹⁰
<i>Pseudomonas aeruginosa</i>	Detected ⁹	Not Detected ¹⁰
<i>Stenotrophomonas maltophilia</i>	Detected ⁹	Not Detected ¹⁰

Yeasts ¹¹				
Pathogen ⁷	LOT	21711150 Control ⁵ 1-GN	LOT	21711151 Control ⁵ 2-GPY
<i>Candida albicans</i>		Not Detected ¹⁰		Detected ⁹
<i>Candida auris</i>		Not Detected ¹⁰		Detected ⁹
<i>Candida glabrata</i>		Not Detected ¹⁰		Detected ⁹
<i>Candida krusei</i>		Not Detected ¹⁰		Detected ⁹
<i>Candida parapsilosis</i>		Not Detected ¹⁰		Detected ⁹
<i>Candida tropicalis</i>		Not Detected ¹⁰		Detected ⁹
<i>Cryptococcus neoformans/gatti</i>		Not Detected ¹⁰		Detected ⁹

1 Quality Control for the BIOFIRE® BCID2 Panel

Contrôle de la qualité du BIOFIRE BCID2 Panel / Qualitätskontrolle für das BIOFIRE BCID2 Panel / Controllo di qualità per il BIOFIRE BCID2 Panel / Control de calidad de BIOFIRE BCID2 Panel / Kvalitetskontroll för BIOFIRE BCID2 Panel

2 Once opened, the product may be used until expiration date.

Une fois ouvert, le produit peut être utilisé jusqu'à la date de date de péremption. / Einmal geöffnet kann das Produkt bis zum angegebenen Ablaufdatum verwendet werden. / Una volta aperto, il prodotto può essere utilizzato fino alla data di scadenza. / Una vez abierto, puede utilizarse el producto hasta la fecha de caducidad. / När produkten har öppnats kan den användas fram till utgångsdatumet.

3 Antimicrobial Resistance Genes

Gènes de résistance aux antimicrobiens / Antimikrobielle Resistenzgene / Geni di resistenza antimicrobica / Genes con resistencia antimicrobiana / Antimikrobiella resistensgener

4 Gene

Gènes / Gen / Gene / Gen / Gen

5 Control

Contrôle / Kontrolle / Controllo / Control / Kontroll

6 Gram-Positive Bacteria

Bactéries Gram-positives / Grampositive Bakterien / Batteri Gram-positivi / Bacterias grampositivas / Grampositiva bakterier

7 Pathogen

Pathogènes / Pathogen / Agente patogeno / Patógeno / Patogen

8 Gram-Negative Bacteria

Bactéries Gram-négatives / Gramnegative Bakterien / Batteri Gram-negativi / Bacterias gramnegativas / Gramnegativa bakterier

9 Detected

Déteçté / Entdeckt / Rilevato / Detectado / Detekterad

10 Not Detected

Non déteçté / Nicht entdeckt / Non rilevato / No detectado / Ej detekterad

11 Yeasts

Levures / Hefen / Lieviti / Levaduras / Jästsvampar

The brand and product names of the instruments are trademarks of their respective holders.

Les noms de marques et de produits des instruments appartiennent à leurs détenteurs respectifs. / Die Marken- und Produktnamen der Geräte sind Warenzeichen der jeweiligen Inhaber. / Le marche e i nomi degli strumenti sono marchi registrati dei rispettivi titolari. / Los nombres de marcas y productos de los instrumentos son marcas comerciales de sus titulares respectivos. / Instrumentmärken och produktnamn är varumärkta av respektive innehavare.