

Dynal MPC™-50

Magnetic Particle Concentrator

For laboratory use only

For optimal results, please read this instruction leaflet before use.

1 PRODUCT DESCRIPTION

- 1.1 Intended Use
- 1.2 Principle of Isolation
- 1.3 Product Description
- 1.4 Technical Support

2. GENERAL INFORMATION

- 2.1 Precautions
- 2.2 Handling
- 2.3 Disinfection
- 2.4 Storage and Stability
- 2.5 Warning and Limitations
- 2.6 Patents and Trademarks
- 2.7 Intellectual Property Disclaimer
- 2.8 Limited Use Label License
- 2.9 Warranty

1. PRODUCT DESCRIPTION

1.1 Intended Use

Dynal MPC™-50 is a magnetic particle concentrator (magnet) for use with standard 50 ml tubes. The use of Dynal MPC-50 allows the separation of Dynabeads® from diverse liquid sample matrices.

1.2 Principle of Isolation

Dynabeads are mixed with the sample in a tube. The Dynabeads will bind to the target material during a short incubation, and then the bead-bound material is separated by the magnet. Dynabeads will be attracted to the side of the tube facing the magnet when the tube is inserted into the Dynal MPC-50. This enables easy removal of supernatant. The Dynabeads with bound material are left in the tube.

1.3 Product Description


The Dynal MPC-50 Magnetic Particle Concentrator contains an ABS plastic housing incorporating high-energy neodymium magnets.

It is designed to hold one standard 50 ml tube with working volumes ranging from 15 ml to 50 ml.

The Dynal MPC-50 is made of disinfectant proof material for easy cleaning (see section 2.2).

Nominal magnetic properties

High performance rare earth magnet alloy, Neodymium Iron Boron (NdFeB).



Residual magnetic flux density (Br) at working distance from the magnet (at the surface of the magnet assembly): 4,500 Gauss (0.45 Tesla)

1.4 Technical Support

Please contact Invitrogen Dynal for further technical information (see contact details).

2. GENERAL INFORMATION

Invitrogen Dynal AS complies with the Quality System Standards ISO 9001:2000 and ISO 13485:2003.

2.1. Precautions

This product contains very strong permanent magnets.

People wearing a pacemaker should not use this product. Pacemakers may be affected or damaged if they come in close contact with a strong magnetic field.

Keep the Dynal MPC-50 magnet away from credit cards and all other products containing magnetic recording devices, such as videocassettes and computer discs. Keep away from delicate instruments, watches, electronic equipment, displays and monitors.

The Health and Safety Officer should take all necessary steps and full responsibility to ensure that the precautions and statements are followed and adhered to. IN NO EVENT SHALL INVITROGEN DYNAL BE LIABLE FOR ANY SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES.

2.2 Handling

Any laboratory skilled in using conventional laboratory techniques that follow good laboratory practices may use Dynal MPC-50 for separation of Dynabeads.

The magnet may attract steel or other magnetic material with high mechanical forces. Great care is needed during handling.

Prevent any form of contact between two magnets! If contact has been made, do not pull apart. Twist off to prevent damage to the unit or fingers.

2.3 Disinfection

The following materials have been tested for cleaning purposes. Spray and/or wipe the Dynal MPC-50 unit with one of the following cleaning agents:

- 70% isopropyl alcohol
- 1% sodium hypochlorite solution (Bleach)
- 0.1N HCl solution

Other disinfectants have not been tested and may not be suitable.

Do not submerge in aqueous solutions. And avoid prolonged exposure to water or aqueous solutions.

Clean with a damp cloth and mild detergent when exposed to harsh solvents.

Do not autoclave the Dynal MPC-50.

2.4 Storage and Stability

Your Dynal MPC-50 contains sintered rare-earth Neodymium Iron Boron permanent magnets. Magnetic strength will not diminish significantly during the lifetime of the product. Loss of magnetic strength will occur only if the magnets are exposed to temperatures greater than 90°C.

Store the product in a cool, dry environment.

Do not leave Dynal MPC-50 exposed to artificial UV light, as the surface material may become brittle.

2.5 Warning and Limitations

The Dynal MPC-50 is only guaranteed to satisfactorily isolate Dynabeads, not the isolation of your specific biomolecule. The efficacy of magnetic separation is – as with all biological techniques – critically dependant on the specificity and avidity of the antibody or other ligand applied, as well as factors concerning the bio-molecules themselves and the matrix from which they are to be isolated.

2.6 Patents and Trademarks

Dynal®, Dynabeads® and Dynal MPC™ are either registered trademarks or trademarks of Invitrogen Dynal AS,

Oslo, Norway. Any registration or trademark symbols used herein denote the registration status of trademarks in the United States. Trademarks may or may not be registered in other countries.

2.7 Intellectual Property Disclaimer

Invitrogen Dynal will not be responsible for violations or patent infringements that may occur with the use of our products.

2.8 Limited Use Label License

No. 5: Invitrogen Technology – The purchase of this product conveys to the buyer the non-transferable right to use the purchased amount of the product and components of the product in research conducted by the buyer (whether the buyer is an academic or for-profit entity). The buyer cannot sell or otherwise transfer (a) this product (b) its components or (c) materials made using this product or its components to a third party or otherwise use this product or its components or materials made using this product or its components for Commercial Purposes. The buyer may transfer information or materials made through the use of this product to a scientific collaborator, provided that such transfer is not for any Commercial Purpose, and that such collaborator agrees in writing (a) not to transfer such materials to any third party, and (b) to use such transferred materials and/or information solely for research and not for Commercial Purposes. Commercial Purposes means any activity by a party for consideration and may include, but is not limited to: (1) use of the product or its components in manufacturing; (2) use of the product or its components to provide a service, information, or data; (3) use of the product or its components for therapeutic, diagnostic or prophylactic purposes; or (4) resale of the product or its components, whether or not such product or its components are resold for use in research. Invitrogen Corporation will not assert a claim against the buyer of infringement of patents owned or controlled by Invitrogen Corporation which cover this product based upon the manufacture, use or sale of a therapeutic, clinical diagnostic, vaccine or prophylactic product developed in research by the buyer in which this product or its components was employed, provided that neither this product nor any of its components was used in the manufacture of such product. If the purchaser is not willing to accept the limitations of this limited use statement, Invitrogen is willing to accept return of the product with a full refund. For information on purchasing a license to this product for purposes other than research, contact

Licensing Department,
Invitrogen Corporation,
1600 Faraday Avenue, Carlsbad,
California 92008.
Phone (760) 603-7200.
Fax (760) 602-6500.
Email: outlicensing@invitrogen.com

2.9 Warranty

The products are warranted to the original purchaser only to conform to the quantity and contents stated on the vial and outer labels for the duration of the stated shelf life. Invitrogen Dynal's obligation and the purchaser's exclusive remedy under this warranty is limited either to replacement, at Invitrogen Dynal's expense, of any products which shall be defective in manufacture, and which shall be returned to Invitrogen Dynal, transportation prepaid, or at Invitrogen Dynal's option, refund of the purchase price.

Claims for merchandise damaged in transit must be submitted to the carrier.

This warranty shall not apply to any products which shall have been altered outside Invitrogen Dynal, nor shall it apply to any products which have been subjected to misuse or mishandling. ALL OTHER WARRANTIES, EXPRESSED, IMPLIED OR STATUTORY, ARE HEREBY SPECIFICALLY EXCLUDED, INCLUDING BUT NOT LIMITED TO WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. Invitrogen Dynal's maximum liability is limited in all events to the price of the products sold by Invitrogen Dynal. IN NO EVENT SHALL INVITROGEN DYNAL BE LIABLE FOR ANY SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES. Some states do not allow limits on warranties, or on remedies for breach in certain transactions. In such states, the limits set forth above may not apply.

Invitrogen Dynal is a part of the Invitrogen Group.

Contact details for your local Invitrogen sales office/technical support can be found at <http://www.invitrogen.com/contact>

© Copyright 2007 Invitrogen Dynal AS, Oslo, Norway.
All rights reserved.

Revised: 02.2007
Printed: 02.2007